

Calibration unit for PGT 120, with 8 DIP switches



- with 8 dip-switches
- to check the resistance limit values on site
- Factory calibrated incl. certificate.
- For later checking of an instrument after a defined calibration interval
- ESD measuring and testing equipment is calibrated in accordance with the EN/ISO/IEC 17025:2005 standard.

item number	WL24341
model	7100.PGT120.CU.8
manufacturer	WARMBIER
manufacturer item number	7100.PGT120.CU.8
order unit	1 piece

content unit

1 piece

