

Calibration unit for PGT 120, with 12 DIP switches



- with 12 dip-switches
- to check the resistance limit values on site
- Factory calibrated incl. certificate.
- For later checking of an instrument after a defined calibration interval
- ESD measuring and testing equipment is calibrated in accordance with the EN/ISO/IEC 17025:2005 standard.

Item no.	WL42975
Model	7100.PGT120.CU.12
Manufacturer	WARMBIER
Manufacturer article no.	7100.PGT120.CU.12
GPSR manufacturer data	Wolfgang WARMBIER Untere Gießwiesen 21

DE-78247 Hilzingen
www.warmbier.com

Sales unit	1 piece
Content unit	1 piece
Incl. battery	no
ESD compliant	yes

